

XCP Octane Analyzer

Next Generation of XCP Automation

CFR Engines Inc. has extended the capabilities of its well-received XCP™ Digital Octane Panel by now offering users an updated version of the fully automated Octane Analyzer (OA) option that significantly reduces operator involvement.

COMPLIANCE

The Octane Analyzer upgrade option unlocks capabilities to test all procedures of the current ASTM Methods:

D2699 – Research Octane Number D2700 – Motor Octane Number

EFFICIENCY

Adding automation to your octane testing delivers the benefits of faster test results, and thus more time to manage additional responsibilities within the lab

ACCURACY

Octane Analyzer reduces operator interaction with the instrument during the test procedure and thus delivers the opportunity for greater reproducibility between sweeps, passes, and full tests

- Automatically complete full 2-pass tests setup test, fill fuel reservoirs, hit start and then wait for the report
- Allows for broader Octane Number ranges compliant with Procedure D 72-103.5 MON and 72-108 RON
- No manual intervention XCP™ Technology with Octane Analyzer option does all fuel switching, fuel/air adjustments, and data recording
- Reduced time for completing passes approximately 10 minutes for single pass, 20 minutes for a full octane determination (2-pass)
- Reduced fuel consumption approximately 100 ml for single sweep



Fully Integrated Solutions by CFR

The Octane Analyzer (OA) upgrade option by CFR builds upon the proven performance of the XCP™ control system. Like all CFR Engines Inc. products, the Octane Analyzer is designed to easily integrate with existing CFR® units and systems.

Whether you are working with a complete unit, an upgrade/ conversion kit, or a genuine CFR service part; you can trust that the product has been designed, manufactured and fully tested to work as an integrated and reliable solution for your operation.

The OA option is offered in two configurations — an add-on upgrade kit for existing XCP or included in complete XCP with OA upgrade kits. Additionally, adding Octane Analyzer to an XCP system unlocks specific OA testing capabilities already built into the XCP Technology platform.

CONFIGURATIONS

Add OA to existing XCP system **p/n: G-802-54** upgrade kit

Add an XCP system with OA to an existing Legacy CFR unit p/n: G-802-53 conversion kit

FLEXIBILITY

Customize Octane testing to your operation. Easily switch between RON or MON, single pass or multiple pass, with falling level or with Octane Analyzer

REQUIREMENTS

- XCP operating on Windows 7
- Sample cooling for unknowns
- Standard unit maintenance

REPORT ACCOUNTABILITY

- Fully Integrated
- Method Compliant Results
- Consistent Format
- LIMS Connectivity
- Complete Data Collection
- Automatic Curve Generation
- Test Event Logging



			REPORT SL	IMMARY			
	DATE	10-12-2018	METHOD	RON	CYLINDER HEIGHT	INCHES	DIGITAL COUNTE
2 FD	SAMPLE	93.4 p3	PROCEDURE	Auto Dy	COMPENSATED	0.44617	798
JFK	Low PRF ON	93	OPERATOR	Ravi	UNCOMPENSATED	0.461331	776
	High PRF ON	95	TIME	12:02:35		in hg.	kpa
					BAROMETER	29.13	98.65
		ACTUAL ON RESULTS	ROUNDED ON RESULT	s			
	SAMPLE ON PASS1	93.18	93.2				
	SAMPLE ON PASS2	93.2	93.2		ON DIFF PASS 2-1	0.0	
	SAMPLE ON PASS3	93.3	93.3		ON DIFF PASS 3-2	0.1	
				_			
	AVERAGE PASS 1&2	93.19	93.2				
	AVERAGE PASS 2&3	93.25	93.3				
TIME	Reference	Fuel	Pass 1 Max KI Level	Max KI	IAT SP	IAT ACTUAL	ON RESULT
						_	UN RESULT
12:02:35	Unknown	93.4 p3	494	64.4	113	112.92	93.18
12:07:06	Ref Fuel ON (PRF)	93	586	65.5	113	113.07	93.18
12:11:37	High Fuel ON (PRF)	95	583	53.6	113	112.75	
DIL PRESSURE PSI	CC VACUUM IN H20	OILTSP	OIL T ACTUAL	WATER TEMP	CONDENSER TEMP	MIXTSP	MIX T ACTUA
28.42	-1.88	130.1	130.24	208.58	84.73	0	
27.64	-2.09	130.1	130.66	208.59	85.09	0	
29.19	-2.06	130.1	129.82	208.66	84.24	0	
			Pass 2				
TIME	0	0	Max KI Level	Max KI	IAT SP	IAT ACTUAL	ON RESULT
12:16:45	Unknown	93.4 p3	497	64.6	113	112.74	
12:21:16	High Fuel ON (PRF)	95	585	53.8	113	113.15	93.2
12:25:46	Ref Fuel ON (PRF)	93	583	65.8	113	113.01	
DIL PRESSURE PSI	CC VACUUM IN H20	OILTSP	OIL T ACTUAL	WATER TEMP	CONDENSER TEMP	MIXTSP	MIX T ACTUA
28.29	-2.06	130.1	130.2	208.6	82.9	0	0
28.16	-2.06	130.1	130.7	208.52	83.69	0	0
27.91	-2.02	130.1	130.02	208.61	83.99	0	0
27.91	-2.02	130.1	130.02	200.01	63.55	0	
			Pass 3				
TIME	0	0	Max KI Level	Max KI	IAT SP	IAT ACTUAL	ON RESULT
12:30:54	Unknown	93.4 p3	501	64.3	113	112.96	93.3
12:35:25	Ref Fuel ON (PRF)	93	587	66.2	113	112.9	
12:39:55	High Fuel ON (PRF)	95	592	53.6	113	113.06	
DIL PRESSURE PSI	CC VACUUM IN H20	OILTSP	OIL T ACTUAL	WATER TEMP	CONDENSER TEMP	MIXTSP	MIX T ACTUA
28.16	-2.16	130.1	130.1	208.56	82.98	0	0
27.77	-2.02	130.1	130.78	208.56	83.69	0	0



